HCl

UNI EN 1911

Sistema manuale per la determinazione dell'HCl in emissione con sonda in titanio (disponibile anche con sonda in vetro).
Norma di riferimento: UNI EN 1911- Emissioni da sorgente fissa - Determinazione della concentrazione in massa di cloruri gassosi espressi come HCl - Metodo di riferimento normalizzato.
Il metodo si applica ai rifiuti gassosi nei quali la concentrazione di cloruri espressa come HCl è compresa tra 1 mg/m³ e 5000 mg/m³ in condizioni normalizzate di temperatura e pressione secondo i valori limite di emissione stabiliti, per esempio, nella Direttiva del Consiglio 2000/76/CE sugli impianti di incenerimento dei rifiuti.

Manual system for the determination of the HCl in emission with titanium probe (available with glass probe).
Reference standard: EN 1911- Emissions from steady/fix source – Determination of the concentration in mass of gaseous chlorides expressed as HCl – Normalized standard method. The method is applied to gaseous wastes in which the chlorides concentration expressed as HCl stands between 1 mg/m³ and 5000 mg/m³ in standard temperature and pressure conditions according to the emission value ranges, as per example, the 2000/76/CE standards on the plants for wastes incineration.